

**TEST FACILITY**

NAMSA  
 115 Chemin de L'Ision  
 38670 Chasse-sur-Rhône  
 France  
 Tel : +33 4 78 07 92 34  
 Fax : +33 4 72 24 08 12

**SPONSOR**

PLATIT AG  
 Dr. Hamid Bolvardi  
 Eichholzstr. 9  
 2545 Selzach / SO  
 Switzerland

CONFIDENTIAL


Test Article Name	AlTiN PVD coated Ti6Al4V plates
Article Identification	Reference: Biocompatible uncoated Ti6Al4V material, Batch: AlTiN

Test	Standard	Study number	Results
ISO MTS cytotoxicity test	ISO 10993-5 (2009)	292764	Pass*
ISO Intracutaneous Study in Rabbits (Two Extracts)	ISO 10993-10 (2010)	292765	Pass*

\*: Conclusions apply only to the test article as received and tested.

Only the final reports are legally acceptable.

Approved by:

  
 Alizé Laffitte  
 Associate Study Director

Date

  
 August 05, 2021